

Certificate of Conformity

in compliance with the EC Directive for the harmonization of the legal regulations of the member states concerning the use of electrical equipment within specific voltage limits (2006/95/EC)

Testing Institute : LGA QualiTest GmbH, Tillystr. 2, 90431 Nuremberg, Germany

Applicant : A. Eberle GmbH & Co. KG
Aalener Str. 30/32, 90441 Nuremberg, Germany

Product, trademark, modell-no. : Measuring system for mains quality
Trademark: a-eberle
PQ-Box 100 3U/4I, PQ-Box 100 4U/4I

Technical Data (Auxiliary supply) : 1N~100 - 240 V / 2~ 100 -400 V, 45 - 65 Hz, 12 W
DC 100 - 220 V, 12 W, 300 V CAT IV

Technical Data (Measuring circuits) : 690 / 400 V 3~ + N + PE or 690 / 400 V 3~ + PEN
600 V CAT III or 300 V CAT IV


Samples of the above mentioned product have been tested and found to be in conformity with EN 61010-1:2001 as shown in the test report (reference no. 817 1224-AT1/1).

This certificate has been established by a body notified to the Member States and Commission of the European Communities according to the provisions of the Low Voltage Directive of December 12th 2006.

This certificate of conformity is the result of testing a sample of the product submitted, in accordance with the provisions of the relevant specific standard. It does not imply an assessment of the whole production and does not permit the use of a LGA QualiTest GmbH conformity mark.

Nuremberg, April 2, 2008

LGA QualiTest GmbH
Testing and Certification Body for Product Safety



Christian Sieber